# **Disposable Sterile Medical Rubber Examination Gloves**

#### - Product name, model and specification

Model specification:Disposable sterile medical rubber examnination

Model: power hemp surface、power smooth surface、power-free hemp sur face、power-free smooth surface; 规格: 6和6以下(特小号)、6.5(小号)、7和7.5(中号)、8和8.5(大号)、9和9以上(特大号)

## ☐ Name, address, production address, contact information and after-sales service unit of the registered person/manufacturer

- 1、Name: Hunan Zhenheyikang Medical Instrument Co.,LTD
- 2. Address: No.6 Building Jingxiang Energy, No.55 Xiaguang east road, Gaoxin district Xiangtan, Hunan
- 3. Production address: NO.6 Building, NO.8 building, 3<sup>rd</sup> and 4<sup>th</sup> floor Jingxiang Energy, No.55 Xiaguang east road, Gaoxin district Xiangtan, Hunan, NO 19 Chutian Road, Yuetang District, Xiangtan City, Hunan
- 4、Contact number: +86 17711627084
- 5、 After-sales service unit: Hunan Zhenheyikang Medical Instrument Co.,LTD
- 三、Manufacturing License NO: Hunan Medical Products Administration Production Licence No. 20150013
- 四、Product Registration Certificate No, Product Technical Requirements No: HNFDA Certified 20212140862
- 五、Structure and composition, scope of application
- Product is made of natural rubber latex. One of the powder models is corn powder. The product was sterilized by ethylene oxide and provided sterile. It is used to test the patient on the hands or fingers of the doctor, or to prevent bites on the doctor's hand.
- 六、六、Use method: open the bag can be used
- 七、Contraindications, precautions, warnings and reminders

Product contraindications: None

Damaged packaging prohibited use, expired products prohibited use;

Disposable use, please destroy after use

Avoid high temperature, humidity, direct sunlight

八、Interpretation of graphics, symbols, abbreviations and other contents used in medical device labels











—次性使用

参考使用说明书

包装破损切勿使用

环氧乙烷灭菌

无菌

#### 九、Installation and use instructions or illustrations

This product does not need to be installed, open the bag can be used.

#### 十、Product maintenance and maintenance methods, special storage, transportation conditions, methods

Products should be stored in relative humidity ≤80%, non-corrosive gas and well-ventilated indoor, ordinary transport, in the transportation, storage process should be kept clean and pollution-free, strictly prohibited rain and snow immersion.

### +-、Date of production, shelf life or expiration date

Valid for two years, production batch number, production date, expiry date please refer to the small package or qualification label.

# +=: Accessories list, including accessories, appendix, lossitage replacement cycle and replacement method description

This product has no accessories

#### 十三、Date and version of preparation or revision of the specification

This specification was revised in February 2023, version 2023.