# **Disposable Surgical Gown**

## Product name, model and specification

Product name: Disposable surgical gown

Model: Standard performance, and high performance

Specification: S (small), M (medium), L (large), XL (one large), XXL (two large), XXXL (three large)

 □. Name, address, production address, contact information and after-sales service unit of the registered person/manufacturer

- 1、Name: Hunan Zhenheyikang Medical Instrument Co.,LTD
- 2. Address: No.6 Building Jingxiang Energy, No.55 Xiaguang east road, Gaoxin district Xiangtan, Hunan
- 3 Production address: NO.6 Building, NO.8 building, 3<sup>rd</sup> and 4<sup>th</sup> floor Jingxiang Energy, No.55 Xiaguang east road, Gaoxin district Xiangtan, Hunan, NO 19 Chutian Road, Yuetang District, Xiangtan City, Hunan
- 4、 Contact number: +86 17711627084
- 5、 After-sales service unit: Hunan Zhenheyikang Medical Instrument Co.,LTD
- 三、Manufacturing License NO: Hunan Medical Products Administration Production Licence No. 20150013
- 四、Product Registration Certificate No, Product Technical Requirements No: HNFDA Certified 20182140024
- 五、Structure and composition, scope of application

The main material of surgical clothing is made of SMS non-woven fabric; the product shall have no skin irritation or delayed hypersensitivity; the product shall be sterile after ethylene oxide sterilization, and the residual amount shall not be greater than  $10 \mu g/g$ .

The product is suitable for use in the operating room of medical institutions to prevent the spread of infection sources between patients and medical staff during the surgical procedures and other invasive examinations.

 $\overrightarrow{n}$  Use method: open the bag can be used

 $\pm$ 、Contraindications, precautions, warnings and reminders

Product contraindications: None

Damaged packaging prohibited use, expired products prohibited use;

Disposable use, please destroy after use

Avoid high temperature, humidity, direct sunlight

## 八、Interpretation of graphics, symbols, abbreviations and other contents used in medical device labels



#### 九、Installation and use instructions or illustrations

This product does not need to be installed, open the bag can be used.

# $+ \mathbf{x}$ Product maintenance and maintenance methods, special storage, transportation conditions, methods

Products should be stored in relative humidity ≤80%, non-corrosive gas and well-ventilated indoor, ordinary transport, in the transportation, storage process should be kept clean and pollution-free, strictly prohibited rain and snow immersion.

# +-, Date of production, shelf life or expiration date

Valid for two years, production batch number, production date, expiry date please refer to the small package or qualification label.

+二、Accessories list, including accessories, appendix, lossitage replacement cycle and replacement method description

This product has n o accessories

+ $\equiv$ , Date and version of preparation or revision of the specification This specification was revised in February 2023, version 2023.